

Nitrile examination & protective gloves



your safety comes first

Examination and protective nitrile gloves made of Nitrile Butadiene Rubber (NBR). The gloves are ambidextrous, latex-free, non-sterile, non-powdered and disposable. Gloves have beaded cuff and textured fingertips.



Healthcare



Industry



Food services



Medical device



Personal Protective Equipment

Advantages of Holík nitrile gloves

- + Latex-free
- + Non-powdered
- + Textured fingertips
- + Complies with EU Regulations
- + Resistance to microorganisms
- + Resistance to some hazardous chemicals
- + Suitable for working with food
- + AQL ≤ 1,5

Comply with

- + EU Regulation for Medical Devices 2017/745, Class I
- + EU Regulation for Personal Protective Equipment 2016/425, Category III
- + ES Regulation for Materials and Articles intended to come into contact with food 1935/2004, article 3

Packaging

- 1 Box = 200 Gloves
- 1 Case = 10 Boxes = 2 000 Gloves
- 1 Pallet = 72 Cases = 144 000 Gloves
- 1 Container = 24 Pallets = 3 456 000 Gloves



200 pcs per box

- + Simplified stock logistics
- + Reduced replacement intervals in the workplace
- + Less packaging waste

Size	S	M	L	XL
Number	6-7	7-8	8-9	9-10

www.holik-international.com

Holík International s.r.o.
Za Dvorem 612 / 763 14 Zlín 12 / Czech Republic

medical@holik-international.cz / +420 577 125 500
CRN: 25322214 / VAT: CZ25322214

Nitrile examination & protective gloves



your safety comes first

Additional information

This product has been tested in accordance with EN ISO 374-1:2016+A1:2018 and achieved following performance levels:

Test chemical	EN ISO 374-1:2016+A1:2018 Performance level	EN ISO 374-4:2019 Degradation
K Sodium Hydroxide 40%	6 (> 480)	-3,7 %
P Hydrogen Peroxide 30%	6 (> 480)	19,7 %
T Formaldehyde 37%	2 (> 30)	0,2 %

Degradation level indicates the change in puncture resistance of the material after continuous contact of the outer surface of the glove with the test chemical. Standard EN ISO 374-4:2019.

Standard EN 16523-1:2015+A1:2018 Degradation levels indicate the change in puncture resistance of the gloves after exposure.

Breakthrough times (min)	> 10	> 30	> 60	> 120	> 240	> 480
Performance level	1	2	3	4	5	6

Standards

EN ISO 21420:2020: Protective gloves - General requirements and test methods

EN ISO 374-1:2016+A1:2018: Protective gloves against hazardous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical hazards

EN ISO 374-2:2019: Protective gloves against hazardous chemicals and micro-organisms - Part 2: Determination of resistance to penetration

EN 16523-1:2015+A1:2018: Determination of resistance to permeation of chemicals - Part 1: Permeation of potentially hazardous liquid chemicals under continuous contact conditions

EN ISO 374-4:2019: Protective gloves against hazardous chemicals and micro-organisms - Part 4: Determination of resistance to degradation by chemicals

EN ISO 374-5:2016: Protective gloves against hazardous chemicals and micro-organisms - Part 5: Terminology and requirements performance requirements for risks due to microorganisms

EN 455-1:2020: Disposable medical gloves - Part 1: Requirements and testing for absence of holes

EN 455-2:2015: Disposable medical gloves - Part 2: Requirements and testing for physical properties

EN 455-3:2015: Disposable medical gloves - Part 3: Requirements and testing for biological evaluation

EN 455-4:2009: Disposable medical gloves - Part 4: Requirements and testing for shelf life determination

Ministry of Health Decree No. 38/2001 Coll. on hygiene requirements for products intended to come into contact with food and foodstuffs.

EN 1186-1:2002 - Materials and articles in contact with foodstuffs - Plastics - Part 1: Guide to the selection of conditions and test methods for overall migration

EN 1186-3:2002 - Materials and Articles in Contact with Foodstuffs - Plastics - Part 3: Test Methods for Overall Migration into Aqueous Food Simulants by Total Immersion

Holík International s.r.o. has set up an integrated management system according to standards:

ČSN EN ISO 13485:2016: Quality management system for medical device manufacturers

ČSN EN ISO 14971:2020: Application of risk management to medical devices

ČSN EN ISO 14001:2016: Environmental management system

ČSN EN ISO 9001:2016: Quality management system

